

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130129</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>JAMES MICHELS</u>	ID# <u>6502</u> Date <u>12/20/10</u>
A Agency <u>NORTH SLOPE BOROUGH P D</u> Phone # <u>907-852-0311</u>	
Instrument Location <u>BARROW / SYND ROOM</u>	
B Alco S/N <u>X301568</u> Target Value <u>.084</u> High Pressure <u>275-300 psi</u>	
Alco Test Values <u>.084</u> <u>.084</u> 1st Alco 2nd Alco	
Signature <u>[Signature]</u> <u>SGT. MICHELS</u>	<u>CB</u> <u>12/30/10</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of MAR, 2011 DN

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129

COB 12/30/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130129

DECEMBER 20, 2010

OPERATOR'S NAME:

NICHOLS/JAMES/D

OPERATOR'S NUMBER: 6502

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: BRWI

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X301568

## --- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.99 in		
ALCO TARGET	.004	10:24
BLANK TEST	.000	10:25
INTERNAL STANDARD	VERIFIED	10:25
ALCO TV 29.99 in	.004	10:25
BLANK TEST	.000	10:26
SUBJECT SAMPLE	.000	10:27
BLANK TEST	.000	10:28
ALCO TV 29.99 in	.004	10:28
BLANK TEST	.000	10:29

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130129

DECEMBER 20, 2010

TIME 10:22

## --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/00/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

## PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ\]^\_`abcdefghijklmno  
pqrstuvwxyz{|}~"